

Your reference

Your message of
2004-06-03Our reference
II.2-460/04Our Phone (direct)
+49 30 8104-3407Our Fax (direct)
+49 30 8104-12 27Our E-Mail
Fachgruppe-II.2@bam.deDate
2004-07-22

HBAM · Federal Institute for Materials Research and Testing · D-12200 Berlin

Mr. Pfeil
Air-be-c Beatmungstechnik GmbH
Wiesestraße 107**07548 Gera****Classification of an article in accordance with the ICAO-TI**


Dear Mr. Pfeil,

to answer your request from 3 June 2004 to classify an article in accordance with the Technical Instructions for the Safe Transport of Dangerous Goods by Air (ICAO-TI) we would like to inform you as follows:

It is the opinion of the BAM that the mentioned article O-zwei® mobil 2.0 which was described by the sent documents is not a dangerous good in accordance with the ICAO-TI if it is not in use during the flight. The reason for that is that there are no dangerous goods within the article. The oxygen will only be given by increasing the concentration by using a molecular filter to take the nitrogen away from the air. There is no pressure receptacle inside. Furthermore it is the opinion of the BAM that the article does not include any other dangerous goods which require a classification in accordance with the ICAO-TI.

The preparation of this certificate is for a charge. You will receive the invoice under separate cover.

Best regards,

**Dipl.-Ing. I. Döring**
Section II.21 "Dangerous Goods/Dangerous Substances;
Assessment and Coordination"Headquarters Lichterfelde:
Unter den Eichen 87
D-12205 BerlinBranch Fabeckstraße:
Unter den Eichen 44-46
D-12203 BerlinBranch Adlershof:
Richard-Willstätter-Straße 11
D-12489 BerlinPhone: +49(0)30/81 04-0
Fax: +49(0)30/8 11 20 29
E-Mail: info@bam.de
Internet: <http://www.bam.de>Bank account:
Bundeskasse Berlin
Deutsche Bundesbank
Filiale Berlin:
Account No. 10 001 018
Sort Code (BLZ) 100 000 00Postbank Berlin:
Account No. 200 102
Sort Code (BLZ) 100-100 10